Radioactive Waste Log

Package ID#	Date into Storage	Tech	Isotopes	Date Released	Background (mR/hr)	Radiation Level at Disposal (mR/hr)	Survey Meter (S/N)	Tech

Hold the radioactive packages for at least 10 T1/2's, then survey in a low background area with all interposed shielding removed. If a package surface mR/hr exceeds background mR/hr, return the package to storage for additional decay.

253.432.2806
www.olympichp.com

Olympic HEALTH PHYSICS

© Olympic Health Physics